



Keramische/Céramiques BRACKETS



Chic. Type Roth*



CHIC. Type Roth MAXILLARY

| Tooth | Torque | Angulation | P/N | | | |
|----------|--------|------------|--------------|--------------|--------------|--------------|
| | | | .018" | | .022" | |
| | | | R | L | R | L |
| U 1 | 12° | 5° | 14-1110-0010 | 14-1210-0010 | 14-2110-0010 | 14-2210-0010 |
| U 2 | 8° | 9° | 14-1120-0010 | 14-1220-0010 | 14-2120-0010 | 14-2220-0010 |
| U 3 | -2° | 13° | 14-1130-0010 | 14-1230-0010 | 14-2130-0010 | 14-2230-0010 |
| U 3/Hk | -2° | 13° | 14-113D-0010 | 14-123D-0010 | 14-213D-0010 | 14-223D-0010 |
| U 4-5 | -7° | 0° | 14-1540-0010 | 14-1540-0010 | 14-2540-0010 | 14-2540-0010 |
| U 4-5/Hk | -7° | 0° | 14-114D-0010 | 14-124D-0010 | 14-214D-0010 | 14-224D-0010 |

MANDIBULAR

| Tooth | Torque | Angulation | P/N | | | |
|--------|--------|------------|--------------|--------------|--------------|--------------|
| | | | .018" | | .022" | |
| | | | R | L | R | L |
| L 1-2 | 0° | 0° | 14-1610-0010 | 14-1610-0010 | 14-2610-0010 | 14-2610-0010 |
| L 3 | -11° | 7° | 14-1430-0010 | 14-1330-0010 | 14-2430-0010 | 14-2330-0010 |
| L 3/Hk | -11° | 7° | 14-143D-0010 | 14-133D-0010 | 14-243D-0010 | 14-233D-0010 |
| L 4 | -17° | 0° | 14-1640-0010 | 14-1640-0010 | 14-2640-0010 | 14-2640-0010 |
| L 4/Hk | -17° | 0° | 14-144D-0010 | 14-134D-0010 | 14-244D-0010 | 14-234D-0010 |
| L 5 | -22° | 0° | 14-1650-0010 | 14-1650-0010 | 14-2650-0010 | 14-2650-0010 |
| L 5/Hk | -22° | 0° | 14-145D-0010 | 14-135D-0010 | 14-245D-0010 | 14-235D-0010 |

Keramische brackets zijn verpakt per 5 stuks /
Les brackets céramiques sont emballés par 5 pcs

GC Orthodontics version of the **MBT™ System** prescription
(McLaughlin, Bennet or Trevisi) and is not a copy of existing system.
No endorsement by the Drs McLaughlin, Bennett or Trevisi is implied.

GC Orthodontics version of the **Roth System** prescription
and is not a copy of existing system.
No endorsement by Dr Roth is implied.





Chic. Type MBT*



CHIC. Type MBT*

MAXILLARY

| Tooth | I/Omm | Torque | Angulation | P/N .018" | | P/N .022" | |
|----------------|-------|--------|------------|--------------|--------------|--------------|--------------|
| | | | | R | L | R | L |
| U 1 | 0.80 | 17° | 4° | 14-1110-0020 | 14-1210-0020 | 14-2110-0020 | 14-2210-0020 |
| U 2 | 1.32 | 10° | 8° | 14-1120-0020 | 14-1220-0020 | 14-2120-0020 | 14-2220-0020 |
| U 3/Hk | 0.73 | -7° | 8° | 14-113D-0020 | 14-123D-0020 | 14-213D-0020 | 14-223D-0020 |
| U 3/Hk | 0.73 | 0° | 8° | 14-113D-0021 | 14-123D-0021 | 14-213D-0021 | 14-223D-0021 |
| U 4-5 | 0.77 | -7° | 0° | 14-1540-0020 | 14-1540-0020 | 14-2540-0020 | 14-2540-0020 |
| U 5/(Optional) | 1.07 | -7° | 0° | 14-1550-0020 | 14-1550-0020 | 14-2550-0020 | 14-2550-0020 |
| U 4-5/Hk | 0.77 | -7° | 0° | 14-114D-0020 | 14-124D-0020 | 14-214D-0020 | 14-224D-0020 |

MANDIBULAR

| Tooth | I/Omm | Torque | Angulation | P/N .018" | | P/N .022" | |
|--------|-------|--------|------------|--------------|--------------|--------------|--------------|
| | | | | R | L | R | L |
| L 1-2 | 1.26 | -6° | 0° | 14-1610-0020 | 14-1610-0020 | 14-2610-0020 | 14-2610-0020 |
| L 3/Hk | 0.83 | -6° | 3° | 14-143D-0020 | 14-133D-0020 | 14-243D-0020 | 14-233D-0020 |
| L 3/Hk | 0.83 | 0° | 3° | 14-143D-0021 | 14-133D-0021 | 14-243D-0021 | 14-233D-0021 |
| L 4 | 0.95 | -12° | 0° | 14-1640-0020 | 14-1640-0020 | 14-2640-0020 | 14-2640-0020 |
| L 4/Hk | 0.95 | -12° | 0° | 14-144D-0020 | 14-134D-0020 | 14-244D-0020 | 14-234D-0020 |
| L 5 | 0.95 | -17° | 0° | 14-1650-0020 | 14-1650-0020 | 14-2650-0020 | 14-2650-0020 |
| L 5/Hk | 0.95 | -17° | 0° | 14-145D-0020 | 14-135D-0020 | 14-245D-0020 | 14-235D-0020 |

Keramische brackets zijn verpakt per 5 stuks / Les brackets céramiques sont emballés par 5 pcs

CHIC. KITS

| Prescription | Upper | Lower | Hook on | P/N | |
|--------------|-------|-------|---------|--------------|--------------|
| | | | | .018" | .022" |
| Roth | 3-3 | — | 3 | 14-1E03-0010 | 14-2E03-0010 |
| Roth | 3-3 | 3-3 | 3 | 14-1EE3-0010 | 14-2EE3-0010 |
| Roth | 5-5 | — | 3 | 14-1C03-0010 | 14-2C03-0010 |
| Roth | 5-5 | 3-3 | 3 | 14-1CE3-0010 | 14-2CE3-0010 |
| Roth | 5-5 | 5-5 | 3 | 14-1CC3-0010 | 14-2CC3-0010 |
| Roth | 5-5 | 3-3 | 3,4,5 | 14-1CE5-0010 | 14-2CE5-0010 |
| Roth | 5-5 | 5-5 | 3,4,5 | 14-1CC5-0010 | 14-2CC5-0010 |
| Type MBT | 3-3 | — | 3 | — | 14-2E03-0020 |
| Type MBT | 5-5 | — | 3 | 14-1C03-0020 | 14-2C03-0020 |
| Type MBT | 5-5 | 3-3 | 3 | — | 14-2CE3-0020 |
| Type MBT | 5-5 | 5-5 | 3 | 14-1CC3-0020 | 14-2CC3-0020 |
| Type MBT | 5-5 | 3-3 | 3,4,5 | — | 14-2CE5-0020 |
| Type MBT | 5-5 | 5-5 | 3,4,5 | 14-1CC5-0020 | 14-2CC5-0020 |



EXPERIENCE ceramic

Type Roth*



EXPERIENCE CERAMIC Type Roth

MAXILLARY

| Tooth | I/O mm | Torque | Angulation | Rotation | P/N | | | |
|----------|--------|--------|------------|----------|--------------|--------------|--------------|--------------|
| | | | | | .018" | | .022" | |
| | | | | | R | L | R | L |
| U 1 | .81 | 12° | 5° | 0° | 11-1110-0010 | 11-1210-0010 | 11-2110-0010 | 11-2210-0010 |
| U 2 | 1.07 | 8° | 9° | 0° | 11-1120-0010 | 11-1220-0010 | 11-2120-0010 | 11-2220-0010 |
| U 3 | .46 | -2° | 13° | 4°m | 11-1130-0010 | 11-1230-0010 | 11-2130-0010 | 11-2230-0010 |
| U 3/Hk | .46 | -2° | 13° | 4°m | 11-113D-0010 | 11-123D-0010 | 11-213D-0010 | 11-223D-0010 |
| U 4-5 | .46 | -7° | 0° | 2°d | 11-1140-0010 | 11-1240-0010 | 11-2140-0010 | 11-2240-0010 |
| U 4-5/Hk | .46 | -7° | 0° | 2°d | 11-114D-0010 | 11-124D-0010 | 11-214D-0010 | 11-224D-0010 |

MANDIBULAR

| Tooth | I/O mm | Torque | Angulation | Rotation | P/N | | | |
|-----------------|--------|--------|------------|----------|--------------|--------------|--------------|--------------|
| | | | | | .018" | | .022" | |
| | | | | | R | L | R | L |
| L 1-2 | 1.17 | -1° | 2° | 0° | 11-1410-0010 | 11-1310-0010 | 11-2410-0010 | 11-2310-0010 |
| L 3 | .56 | -11° | 7° | 2°m | 11-1430-0010 | 11-1330-0010 | 11-2430-0010 | 11-2330-0010 |
| L 3/Hk | .56 | -11° | 7° | 2°m | 11-143D-0010 | 11-133D-0010 | 11-243D-0010 | 11-233D-0010 |
| L 4(zirconium) | .38 | -17° | 0° | 4°d | 11-1440-Z010 | 11-1340-Z010 | 11-2440-Z010 | 11-2340-Z010 |
| L 4/Hk/ (zirc.) | .38 | -17° | 0° | 4°d | 11-144D-Z010 | 11-134D-Z010 | 11-244D-Z010 | 11-234D-Z010 |
| L 5(zirconium) | .28 | -22° | 0° | 4°d | 11-1450-Z010 | 11-1350-Z010 | 11-2450-Z010 | 11-2350-Z010 |
| L 5/Hk(zirc.) | .28 | -22° | 0° | 4°d | 11-145D-Z010 | 11-135D-Z010 | 11-245D-Z010 | 11-235D-Z010 |

EXPERIENCE CERAMIC Type Roth .020"x .028"

Features an original slot size on upper and lower incisors: .020" x .028". Used in conjunction with .020"x .025" stainless steel archwires, the system controls the anterior torque and brings more rigidity to the buccal segments during space closure.

MAXILLARY


| Tooth | I/O mm | Torque | Angulation | Rotation | P/N | |
|------------------------|--------|--------|------------|----------|--------------|--------------|
| | | | | | R | L |
| U 1 .020" x .028" | .81 | 12° | 5° | 0° | 11-5110-0010 | 11-5210-0010 |
| U 2 .020" x .028" | 1.07 | 8° | 9° | 0° | 11-5120-0010 | 11-5220-0010 |
| U 3 .022" x .028"/Hk | .46 | 0° | 8° | 2°m | 11-213D-0030 | 11-223D-0030 |
| U 4/5 .022" x .028" | .46 | -7° | 0° | 2°d | 11-2140-0010 | 11-2240-0010 |
| U 4/5 .022" x .028"/Hk | .46 | -7° | 0° | 2°d | 11-214D-0010 | 11-224D-0010 |

MANDIBULAR

| Tooth | I/O mm | Torque | Angulation | Rotation | P/N | |
|-------------------------------|--------|--------|------------|----------|--------------|--------------|
| | | | | | R | L |
| L 1/2 .020" x .028" | 1.17 | 0° | 2° | 0° | 11-5410-0010 | 11-5310-0010 |
| L 3 .022" x .028"/Hk | .56 | -3° | 7° | 2°m | 11-243D-0030 | 11-233D-0030 |
| L 4* .022" x .028"/Hk/(zirc.) | .38 | -17° | 0° | 4°d | 11-244D-Z010 | 11-234D-Z010 |
| L 5* .022" x .028"/Hk/(zirc.) | .28 | -22° | 0° | 4°d | 11-245D-Z010 | 11-235D-Z010 |

Keramische brackets zijn verpakt per 5 stuks /
Les brackets céramiques sont emballés par 5 pcs

Keramische/Céramiques BRACKETS



EXPERIENCE ceramic

Extra Torque



EXPERIENCE CERAMIC Extra Torque

This prescription is based on the use of less than full size archwires due to increased torque on the maxillary incisors. Increased buccal root torque in the upper bicuspids prevents any lingual cusp interference.

MAXILLARY

| Tooth | I/O mm | Torque | Angulation | Rotation | P/N | | | |
|----------|--------|--------|------------|----------|--------------|--------------|--------------|--------------|
| | | | | | .018" | | .022" | |
| | | | | | R | L | R | L |
| U 1 | .81 | 18° | 5° | 0° | 11-1110-0030 | 11-1210-0030 | 11-2110-0030 | 11-2210-0030 |
| U 2 | 1.07 | 10° | 8° | 0° | 11-1120-0030 | 11-1220-0030 | 11-2120-0030 | 11-2220-0030 |
| U 3 | .46 | 0° | 8° | 2°m | 11-1130-0030 | 11-1230-0030 | 11-2130-0030 | 11-2230-0030 |
| U 3/Hk | .46 | 0° | 8° | 2°m | 11-113D-0030 | 11-123D-0030 | 11-213D-0030 | 11-223D-0030 |
| U 4-5 | .46 | -10° | 4° | 2°d | 11-1140-0030 | 11-1240-0030 | 11-2140-0030 | 11-2240-0030 |
| U 4-5/Hk | .46 | -10° | 4° | 2°d | 11-114D-0030 | 11-124D-0030 | 11-214D-0030 | 11-224D-0030 |

MANDIBULAR

| Tooth | I/O mm | Torque | Angulation | Rotation | P/N | | | |
|-----------------|--------|--------|------------|----------|--------------|--------------|--------------|--------------|
| | | | | | .018" | | .022" | |
| | | | | | R | L | R | L |
| L 1-2 | 1.17 | 1° | 0° | 0° | 11-1610-0030 | 11-1610-0030 | 11-2610-0030 | 11-2610-0030 |
| L 3 | .56 | -3° | 7° | 2°m | 11-1430-0030 | 11-1330-0030 | 11-2430-0030 | 11-2330-0030 |
| L 3/Hk | .56 | -3° | 7° | 2°m | 11-143D-0030 | 11-133D-0030 | 11-243D-0030 | 11-233D-0030 |
| L 4*/(Exp M) | .33 | -7° | 0° | 2°d | 20-1440-0030 | 20-1340-0030 | 20-2440-0030 | 20-2340-0030 |
| L 4*/Hk/(Exp M) | .33 | -7° | 0° | 2°d | 20-144D-0030 | 20-134D-0030 | 20-244D-0030 | 20-234D-0030 |
| L 5*/(Exp M) | .33 | -7° | 0° | 2°d | 20-1450-0030 | 20-1350-0030 | 20-2450-0030 | 20-2350-0030 |
| L 5*/Hk/(Exp M) | .33 | -7° | 0° | 2°d | 20-145D-0030 | 20-135D-0030 | 20-245D-0030 | 20-235D-0030 |

* Since ceramic material is much harder than enamel, we recommend using EXPERIENCE metal brackets on mandibular bicuspids (L4, L5)

Keramische brackets zijn verpakt per 5 stuks /
Les brackets céramiques sont emballés par 5 pcs





EXPERIENCE ceramic

Type MBT*



EXPERIENCE CERAMIC Type MBT*

MAXILLARY

| Tooth | I/O mm | Torque | Angulation | Rotation | P/N | | | |
|----------------|--------|--------|------------|----------|--------------|--------------|--------------|--------------|
| | | | | | .018" | | .022" | |
| | | | | | R | L | R | L |
| U 1 | .81 | 17° | 4° | 0° | 11-1110-0020 | 11-1210-0020 | 11-2110-0020 | 11-2210-0020 |
| U 2 | 1.07 | 10° | 8° | 0° | 11-1120-0020 | 11-1220-0020 | 11-2120-0020 | 11-2220-0020 |
| U 3/Hk | .46 | -7° | 8° | 0° | 11-113D-0020 | 11-123D-0020 | 11-213D-0020 | 11-223D-0020 |
| U 3/Hk | .46 | 0° | 8° | 0° | 11-113D-0021 | 11-123D-0021 | 11-213D-0021 | 11-223D-0021 |
| U 4-5 | .46 | -7° | 0° | 0° | 11-1540-0020 | 11-1540-0020 | 11-2540-0020 | 11-2540-0020 |
| U 4-5/Hk | .46 | -7° | 0° | 0° | 11-114D-0020 | 11-124D-0020 | 11-214D-0020 | 11-224D-0020 |
| U 5/(Optional) | .96 | -7° | 0° | 0° | 11-1550-0020 | 11-1550-0020 | 11-2550-0020 | 11-2550-0020 |
| U 5/Hk/(Opt) | .96 | -7° | 0° | 0° | 11-115D-0020 | 11-125D-0020 | 11-215D-0020 | 11-225D-0020 |

MANDIBULAR

| Tooth | I/O mm | Torque | Angulation | Rotation | P/N | | | |
|--------|--------|--------|------------|----------|--------------|--------------|--------------|--------------|
| | | | | | .018" | | .022" | |
| | | | | | R | L | R | L |
| L 1-2 | 1.17 | -6° | 0° | 0° | 11-1610-0020 | 11-1610-0020 | 11-2610-0020 | 11-2610-0020 |
| L 3/Hk | .56 | -6° | 3° | 0° | 11-143D-0020 | 11-133D-0020 | 11-243D-0020 | 11-233D-0020 |
| L 3/Hk | .56 | 0° | 3° | 0° | 11-143D-0021 | 11-133D-0021 | 11-243D-0021 | 11-233D-0021 |
| L 4 | .33 | -12° | 0° | 0° | 11-1440-0020 | 11-1340-0020 | 11-2440-0020 | 11-2340-0020 |
| L 4/Hk | .33 | -12° | 0° | 0° | 11-144D-0020 | 11-134D-0020 | 11-244D-0020 | 11-234D-0020 |
| L 5 | .33 | -17° | 0° | 0° | 11-1450-0020 | 11-1350-0020 | 11-2450-0020 | 11-2350-0020 |
| L 5/Hk | .33 | -17° | 0° | 0° | 11-145D-0020 | 11-135D-0020 | 11-245D-0020 | 11-235D-0020 |

Keramische brackets zijn verpakt per 5 stuks /
Les brackets céramiques sont emballés par 5 pcs





EXPERIENCE ceramic Kits



EXPERIENCE ceramic Kits

| | | | | P/N | |
|-----------------------------|------------|------------|-----------|---------------|--------------|
| Prescription | Upper arch | Lower arch | With Hook | .020" x .028" | |
| Type Roth .020" x .028" | 5-5 | 3-3 | 3 | 11-5CE3-0010 | |
| Type Roth .020" x .028" | 5-5 | — | 3 | 11-5C03-0010 | |
| | | | | P/N | |
| Prescription | Upper arch | Lower arch | With Hook | .018" | .022" |
| Type MBT* | 3-3 | — | 3 | 11-1E03-0020 | 11-2E03-0020 |
| Type MBT* | 3-3 | 3-3 | 3 | 11-1EE3-0020 | 11-2EE3-0020 |
| Type MBT* | 5-5 | — | 3 | 11-1C03-0020 | 11-2C03-0020 |
| Type MBT* | 5-5 | — | 3,4,5 | 11-1C05-0020 | 11-2C05-0020 |
| Type MBT* | 5-5 | 3-3 | 3 | 11-1CE3-0020 | 11-2CE3-0020 |
| Type MBT* | 5-5 | 3-3 | 3,4,5 | 11-1CE5-0020 | 11-2CE5-0020 |
| Type MBT* | 5-5 | 5-5 | 3 | 11-1CC3-0020 | 11-2CC3-0020 |
| Type MBT* | 5-5 | 5-5 | 3,4,5 | 11-1CC5-0020 | 11-2CC5-0020 |
| | | | | P/N | |
| Prescription | Upper arch | Lower arch | With Hook | .018" | .022" |
| Type Roth | 3-3 | — | 3 | 11-1E03-0010 | 11-2E03-0010 |
| Type Roth | 3-3 | 3-3 | 3 | 11-1EE3-0010 | 11-2EE3-0010 |
| Type Roth | 5-5 | — | 3 | 11-1C03-0010 | 11-2C03-0010 |
| Type Roth | 5-5 | — | 3,4,5 | 11-1C05-0010 | 11-2C05-0010 |
| Type Roth | 5-5 | 3-3 | 3 | 11-1CE3-0010 | 11-2CE3-0010 |
| Type Roth | 5-5 | 3-3 | 3,4,5 | 11-1CE5-0010 | 11-2CE5-0010 |
| Type Roth (L 4-5 zirconium) | 5-5 | 5-5 | 3 | 11-1CC3-M010 | 11-2CC3-M010 |
| Type Roth (L 4-5 zirconium) | 5-5 | 5-5 | 3,4,5 | 11-1CC5-M010 | 11-2CC5-M010 |
| | | | | P/N | |
| Prescription | Upper arch | Lower arch | With Hook | .018" | .022" |
| Extra Torque | 3-3 | — | 3 | 11-1E03-0030 | 11-2E03-0030 |
| Extra Torque | 3-3 | 3-3 | 3 | 11-1EE3-0030 | 11-2EE3-0030 |
| Extra Torque | 5-5 | — | 3 | 11-1C03-0030 | 11-2C03-0030 |
| Extra Torque | 5-5 | — | 3,4,5 | 11-1C05-0030 | 11-2C05-0030 |
| Extra Torque | 5-5 | 3-3 | 3 | 11-1CE3-0030 | 11-2CE3-0030 |
| Extra Torque | 5-5 | 3-3 | 3,4,5 | 11-1CE5-0030 | 11-2CE5-0030 |
| Prescription | Upper arch | Lower arch | With Hook | .022" | |
| CCO | 5-5 | - | 3,4,5 | 11-2C05-0007 | |
| CCO | 5-5 | 3-3 | 3,4,5 | 11-2CE5-0007 | |



EXPERIENCE ceramic

CCO

EXPERIENCE CERAMIC CCO MAXILLARY

| Tooth | I/O mm | Torque | Angulation | P/N .022" x.028" | |
|----------|--------|--------|------------|---------------------|--------------|
| | | | | R | L |
| U 1 | .81 | 12° | 5° | 11-2110-0007 | 11-2210-0007 |
| U 2 | 1.07 | 10° | 9° | 11-2120-0007 | 11-2220-0007 |
| U 3/Hk | .46 | -7° | 10° | 11-213D-0007 | 11-223D-0007 |
| U 4-5 | .46 | -9° | 0° | 11-2540-0007 | 11-2540-0007 |
| U 4-5/Hk | .46 | -9° | 0° | 11-214D-0007 | 11-224D-0007 |

MANDIBULAR

| Tooth | I/O mm | Torque | Angulation | P/N .022" x.028" | |
|--------|--------|--------|------------|---------------------|--------------|
| | | | | R | L |
| L 1/2 | 1.17 | -6° | 0° | 11-2610-0007 | 11-2610-0007 |
| L 3/Hk | .56 | -8° | 3° | 11-243D-0007 | 11-233D-0007 |
| L 4 | .38 | -12° | 0° | 11-2440-0020 | 11-2340-0020 |
| L 4/Hk | .38 | -12° | 0° | 11-244D-0020 | 11-234D-0020 |
| L 5 | .38 | -17° | 0° | 11-2450-0020 | 11-2350-0020 |
| L 5/Hk | .38 | -17° | 0° | 11-245D-0020 | 11-235D-0020 |



EXPERIENCE ceramic

Optional



OPTIONAL BRACKETS

| Tooth | Torque | Angulation | Rotation | P/N | | | |
|----------|--------|------------|----------|--------------|--------------|--------------|--------------|
| | | | | .018" | .022" | .022" | .022" |
| U 3/Hk | 7° | 5° | 0° | 11-113D-0027 | 11-123D-0027 | 11-213D-0027 | 11-223D-0027 |
| U 4 Univ | 0° | 0° | 0° | 11-1540-0025 | 11-1540-0025 | 11-2540-0025 | 11-2540-0025 |

Keramische brackets zijn verpakt per 5 stuks / Les brackets céramiques sont emballés par 5 pcs

Instruments for working with EXPERIENCE brackets:



C-Tool: Features both a double forked end to facilitate wire insertion and an end with a hollow ball to safely open the bracket from the gingival side.

P/N. **50-1030-0000**



Open Tool: Is a double ended instrument that features a hollow ball, which aids in opening EXPERIENCE brackets from either the buccal (hook) or gingival (safety) side.

P/N. **50-0030-0000**



20%
smaller than
the Experience
Ceramic

EXPERIENCE mini ceramic Type MBT* en Roth*



EXPERIENCE mini Ceramic Type MBT

MAXILLARY

| Tooth | Torque | Angulation | .018 | | .022 | |
|----------------|--------|------------|--------------|--------------|--------------|--------------|
| | | | R | L | R | L |
| U1 | 17°T | 4°A | 12-1110-0020 | 12-1210-0020 | 12-2110-0020 | 12-2210-0020 |
| U2 | 10°T | 8°A | 12-1120-0020 | 12-1220-0020 | 12-2120-0020 | 12-2220-0020 |
| U3 with hook | 0°T | 8°A | 12-113D-0021 | 12-123D-0021 | 12-213D-0021 | 12-223D-0021 |
| U4 univ | -7°T | 0°A | 12-1540-0020 | 12-1540-0020 | 12-2540-0020 | 12-2540-0020 |
| U4&5 with hook | -7°T | 0°A | 12-114D-0020 | 12-124D-0020 | 12-214D-0020 | 12-224D-0020 |

MANDIBULAR

| Tooth | Torque | Angulation | .018 | | .022 | |
|--------------|--------|------------|--------------|--------------|--------------|--------------|
| | | | R | L | R | L |
| L1-2 | -6°T | 0°A | 12-1610-0020 | 12-1610-0020 | 12-2610-0020 | 12-2610-0020 |
| L3 with hook | 0°T | 3°A | 12-143D-0021 | 12-133D-0021 | 12-243D-0021 | 12-233D-0021 |
| L4 | -12°T | 2°A | 12-1440-0020 | 12-1340-0020 | 12-2440-0020 | 12-2340-0020 |
| L4 with hook | -12°T | 2°A | 12-144D-0020 | 12-134D-0020 | 12-244D-0020 | 12-234D-0020 |
| L5 | -17°T | 2°A | 12-1450-0020 | 12-1350-0020 | 12-2450-0020 | 12-2350-0020 |
| L5 with hook | -17°T | 2°A | 12-145D-0020 | 12-135D-0020 | 12-245D-0020 | 12-235D-0020 |

Keramische brackets zijn verpakt per 5 stuks / Les brackets céramiques sont emballés par 5 pcs

EXPERIENCE mini Ceramic Type Roth

MAXILLARY

| Tooth | I/O mm | Torque | Ang. | Rotation | .018 | | .022 | |
|-----------|--------|--------|------|----------|--------------|--------------|--------------|--------------|
| | | | | | R | L | R | L |
| U1 | .81 | 12°T | 5°A | 0° | 12-1110-0010 | 12-1210-0010 | 12-2110-0010 | 12-2210-0010 |
| U2 | 1.07 | 8°T | 9°A | 0° | 12-1120-0010 | 12-1220-0010 | 12-2120-0010 | 12-2220-0010 |
| U3/hook | .46 | -2°T | 13°A | 4°m | 12-113D-0010 | 12-123D-0010 | 12-213D-0010 | 12-223D-0010 |
| U4-5/hook | .46 | -7°T | 0°A | 2°d | 12-114D-0010 | 12-124D-0010 | 12-214D-0010 | 12-224D-0010 |

MANDIBULAR

| Tooth | I/O mm | Torque | Ang. | Rotation | .018 | | .022 | |
|---------|--------|--------|------|----------|--------------|--------------|--------------|--------------|
| | | | | | R | L | R | L |
| L1-2 | 1.17 | -1°T | 2°A | 0° | 12-1410-0010 | 12-1310-0010 | 12-2410-0010 | 12-2310-0010 |
| L3/hook | .56 | -11°T | 7°A | 2°m | 12-143D-0010 | 12-133D-0010 | 12-243D-0010 | 12-233D-0010 |
| L4/hook | .38 | -17°T | 0°A | 4°d | 12-144D-0010 | 12-134D-0010 | 12-244D-0010 | 12-234D-0010 |
| L5/hook | .28 | -22°T | 0°A | 4°d | 12-145D-0010 | 12-135D-0010 | 12-245D-0010 | 12-235D-0010 |

KITS mini Ceramic

| Prescription | Upper | Lower | Hook on | .018 | .022 |
|--------------|-------|-------|---------|--------------|--------------|
| Type Roth | 5-5 | 5-5 | 3,4,5 | 12-1CC5-0010 | 12-2CC5-0010 |
| Type MBT | 5-5 | 5-5 | 3,4,5 | 12-1CC5-0020 | 12-2CC5-0020 |



AGILITY ceramic Type Roth*

AGILITY CERAMIC Type Roth Sold in packages of 5

MAXILLARY Roth¹ Rx

| Tooth | Torque | Angle | Rotation | M/D | R/L | ID | .018 | .022 |
|---|--------|-------|----------|--------|-----|----|--------------|--------------|
| Central | 12° | 5° | 0° | 3.9 mm | R | ● | 127-U1-8R-V | 127-U1-2R-V |
| | | | | | L | ● | 127-U1-8L-V | 127-U1-2L-V |
| Lateral | 8° | 9° | 0° | 3.3 mm | R | ○ | 127-U2-8R-V | 127-U2-2R-V |
| | | | | | L | ○ | 127-U2-8L-V | 127-U2-2L-V |
| Cuspid (without Hook) | 0° | 11° | 0° | 3.4 mm | R | ● | 127-U3-8R-V | 127-U3-2R-V |
| | | | | | L | ● | 127-U3-8L-V | 127-U3-2L-V |
| Cuspid Hk (with Hook) | 0° | 11° | 0° | 3.4 mm | R | ● | 127-U3-8RK-V | 127-U3-2RK-V |
| | | | | | L | ● | 127-U3-8LK-V | 127-U3-2LK-V |
| 1 st & 2 nd Bicuspid | -7° | 0° | 0° | 3.4 mm | U | ●● | 127-U4-8U-V | 127-U4-2U-V |
| 1 st & 2 nd Bicuspid Hk | -7° | 0° | 0° | 3.4 mm | R | ● | 127-U4-8RK-V | 127-U4-2RK-V |
| | | | | | L | ● | 127-U4-8LK-V | 127-U4-2LK-V |

MANDIBULAR Roth¹ Rx

| Tooth | Torque | Angle | Rotation | M/D | R/L | ID | .018 | .022 |
|-----------------------------|--------|-------|----------|--------|-----|----|--------------|--------------|
| Anterior | 0° | 0° | 0° | 2.6 mm | U | ●● | 127-L1-8U-V | 127-L1-2U-V |
| Cuspid (without Hook) | -11° | 5° | 0° | 3.4 mm | R | ● | 127-L3-8R-V | 127-L3-2R-V |
| | | | | | L | ● | 127-L3-8L-V | 127-L3-2L-V |
| Cuspid Hk (with Hook) | -11° | 5° | 0° | 3.4 mm | R | ● | 127-L3-8RK-V | 127-L3-2RK-V |
| | | | | | L | ● | 127-L3-8LK-V | 127-L3-2LK-V |
| 1 st Bicuspid | -17° | 0° | 0° | 3.4 mm | U | ●● | 127-L4-8U-V | 127-L4-2U-V |
| 1 st Bicuspid Hk | -17° | 0° | 0° | 3.4 mm | R | ● | 127-L4-8RK-V | 127-L4-2RK-V |
| | | | | | L | ● | 127-L4-8LK-V | 127-L4-2LK-V |
| 2 nd Bicuspid | -22° | 0° | 0° | 3.4 mm | U | ●● | 127-L5-8U-V | 127-L5-2U-V |
| 2 nd Bicuspid Hk | -22° | 0° | 0° | 3.4 mm | R | ● | 127-L5-8RK-V | 127-L5-2RK-V |
| | | | | | L | ● | 127-L5-8LK-V | 127-L5-2LK-V |

BRACKET KITS Roth¹ Rx


| 20 brackets per kit | .018 | .022 |
|-----------------------------------|---------------|---------------|
| Upper/Lower 5x5 Hook on 3 | 19K-3-18-01 | 19K-3-22-01 |
| Upper/Lower 5x5 Hook on 3, 4, & 5 | 19K-345-18-01 | 19K-345-22-01 |

Brackets are packaged with 1 case oriented in a kit box.

Brackets are also available in kits of 5 or 10 cases oriented in a carousel dispenser.



Agility™ Twin Bracket kit, 20 brackets oriented in a kit box



**MBT is a trademark of 3M Unitek Corporation.

G&H® prescriptions are not implied to be an exact version of any other system, nor do we claim any endorsement by the doctor.



AGILITY ceramic Type MBT*

AGILITY CERAMIC Type MBT Sold in packages of 5

MAXILLARY MBT² Rx

| Tooth | Torque | Angle | Rotation | M/D | R/L | ID | .018 | .022 |
|---|--------|-------|----------|--------|-----|----|---------------|---------------|
| Central | 17° | 4° | 0° | 3.9 mm | R | ● | 128-U1-8R-V | 128-U1-2R-V |
| | | | | | L | ● | 128-U1-8L-V 1 | 28-U1-2L-V |
| Lateral | 10° | 8° | 0° | 3.3 mm | R | ○ | 128-U2-8R-V | 128-U2-2R-V |
| | | | | | L | ○ | 128-U2-8L-V | 128-U2-2L-V |
| Cuspid (without Hook) | -7° | 8° | 0° | 3.4 mm | R | ● | 128-U3-8R-V | 128-U3-2R-V |
| | | | | | L | ● | 128-U3-8L-V | 128-U3-2L-V |
| Cuspid Hk (with Hook) | -7° | 8° | 0° | 3.4 mm | R | ● | 128-U3-8RK-V | 128-U3-2RK-V |
| | | | | | L | ● | 128-U3-8LK-V | 128-U3-2LK-V |
| Cuspid (without Hook) | 0° | 8° | 0° | 3.4 mm | R | ● | 128-U3-8R0-V | 128-U3-2R0-V |
| | | | | | L | ● | 128-U3-8L0-V | 128-U3-2L0-V |
| Cuspid Hk (with Hook) | 0° | 8° | 0° | 3.4 mm | R | ● | 128-U3-8RK0-V | 128-U3-2RK0-V |
| | | | | | L | ● | 128-U3-8LK0-V | 128-U3-2LK0-V |
| 1 st & 2 nd Bicuspid | -7° | 0° | 0° | 3.4 mm | U | ●● | 128-U4-8U-V | 128-U4-2U-V |
| 1 st & 2 nd Bicuspid Hk | -7° | 0° | 0° | 3.4 mm | R | ● | 128-U4-8RK-V | 128-U4-2RK-V |
| | | | | | L | ● | 128-U4-8LK-V | 128-U4-2LK-V |

MANDIBULAR MBT² Rx

| Tooth | Torque | Angle | Rotation | M/D | R/L | ID | .018 | .022 |
|-----------------------------|----------|-------|----------|--------|--------|----|----------------|---------------|
| 128-L1-2U-V | Anterior | -6° | 0° | 0° | 2.6 mm | U | ●● | 128-L1-8U-V |
| Cuspid (without Hook) | -6° | 3° | 0° | 3.4 mm | R | ● | 128-L3-8R-V | 128-L3-2R-V |
| | | | | | L | ● | 128-L3-8L-V | 128-L3-2L-V |
| Cuspid Hk (with Hook) | -6° | 3° | 0° | 3.4 mm | R | ● | 128-L3-8RK-V | 128-L3-2RK-V |
| | | | | | L | ● | 128-L3-8LK-V | 128-L3-2LK-V |
| Cuspid (without Hook) | 0° | 3° | 0° | 3.4 mm | R | ● | 128-L3-8R0-V | 128-L3-2R0-V |
| | | | | | L | ● | 128-L3-8L0-V | 128-L3-2L0-V |
| Cuspid Hk (with Hook) | 0° | 3° | 0° | 3.4 mm | R | ● | 128-L3-8RK0-V | 128-L3-2RK0-V |
| | | | | | L | ● | 128-L3-8LK0-V | 128-L3-2LK0-V |
| 1 st Bicuspid | -12° | 0° | 0° | 3.4 mm | U | ●● | 128-L4-8U-V | 128-L4-2U-V |
| 1 st Bicuspid Hk | -12° | 0° | 0° | 3.4 mm | R | ● | 128-L4-8RK-V | 128-L4-2RK-V |
| | | | | | L | ● | 128-L4-8LK-V | 128-L4-2LK-V |
| 2 nd Bicuspid | -17° | 0° | 0° | 3.4 mm | U | ●● | 128-L5-8U-V | 128-L5-2U-V |
| 2 nd Bicuspid Hk | -17° | 0° | 0° | 3.4 mm | R | ● | 128-L5-8RK-V | 128-L5-2RK-V |
| | | | | | L | ● | 128-L5-8LK-V 1 | 28-L5-2LK-V |

BRACKET KITS MBT² Rx

| 20 brackets per kit | .018 | .022 |
|-----------------------------------|---------------|---------------|
| Upper/Lower 5x5 Hook on 3 | 21K-3-18-01 | 21K-3-22-011 |
| Upper/Lower 5x5 Hook on 3, 4, & 5 | 21K-345-18-01 | 21K-345-22-01 |

MBT Rx kits include 0° torque on the cuspids - custom kits available upon request.

Brackets are packaged with 1 case oriented in a kit box.

Brackets are also available in kits of 5 or 10 cases oriented in a carousel dispenser.



Agility™ Twin Bracket
kit, 20 brackets
oriented in a kit box

